



EU-type examination certificate

Number **T10947** revision 1
Project number 1901121
Page 1 of 1

| | |
|----------------------|--|
| Issued by | NMi Certin B.V., designated and notified by the Netherlands to perform tasks with respect to conformity modules mentioned in article 17 of Directive 2014/32/EU, after having established that the Measuring instrument meets the applicable requirements of Directive 2014/32/EU, to: |
| Manufacturer | Flow Meter Group B.V. Menisstraat 5c 7091 ZZ Dinxperlo The Netherlands |
| Measuring instrument | An Ultrasonic Gas Meter Type : FMU Destined for the measurement of : Gas volume Accuracy class : Class 1,0 Environment classes : M1 / E2 Ambient temperature range : -25 °C / +55 °C Designed for : Condensing humidity Intended location : Open Further properties are described in the annexes: – Description T10947 revision 1; – Documentation folder T10947-3. |
| Valid until | 26 January 2027 |
| Remarks | This revision replaces the earlier versions, including its documentation folder. |

Issuing Authority **NMi Certin B.V., Notified Body number 0122**
16 November 2017



C. Oosterman
Head Certification Board

NMi Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the manufacturer shall indemnify third-party liability.

The designation of NMi Certin BV.as Notified Body can be verified at <http://ec.europa.eu/growth/tools-databases/nando/>

Reproduction of the complete document only is permitted.

