

Issued by NMI Certin B.V.,
designated and notified by the Netherlands to perform tasks with respect to
conformity modules mentioned in article 17 of Directive 2014/32/EU, after
having established that the Measuring instrument meets the applicable
requirements of Directive 2014/32/EU, to:

Manufacturer Beijing United Gas Meters Co., Ltd.
No. 9 Ganluyuan Shilipu Chaoyang District
100025 Beijing
P.R. China

Measuring instrument **A Diaphragm Gas Meter**

Type : Cubix U6 or G4U

Manufacturer's mark or name : UGM

Destined for the measurement of : Gas volume

Accuracy class : Class 1,5

Environment classes : M1 / E1

Temperature range : -10 °C / +40 °C

Location : Closed

G-value	Q _{max} [m ³ /h]	Q _{min} [m ³ /h]	Q _t [m ³ /h]	V [dm ³]
4	6	0,025	0,6	2

Further properties are described in the annexes:
– Description T11224 revision 1;
– Documentation folder T11224-1.

Valid until 18 December 2027

Remarks This revision replaces the earlier version, excluding its documentation folder.

Issuing Authority **NMI Certin B.V., Notified Body number 0122**
1 November 2018



C. Oosterman
Head Certification Board