

EU-type examination certificate

Number **T10159** revision 7
Project number 3936816
Page 1 of 1

Issued by

NMi Certin B.V.,
designated and notified by the Netherlands to perform tasks with respect to
conformity assessment procedures mentioned in article 17 of Directive
2014/32/EU, after having established that the measuring instrument meets
the applicable requirements of Directive 2014/32/EU, to:

Manufacturer

Ex~i Flow Measurement Ltd.
Unit 22, Ford Lane Business Park, Ford Lane
FORD, West Sussex BN18 0UZ
United Kingdom

Measuring instrument

An **electronic gas-volume conversion device (EVCD)**, intended to be
used for gas volume conversion as a sub-assembly (according to article 4 of
the MID) of a gas meter.

Type : SFC3000

Manufacturer's mark or name : Ex~i Flow Measurement

Conversion principle : PTZ

Ambient temperature range : -25 °C / +55 °C

Designed for : non-condensing humidity

Environment classes : M2 / E2

The intended location for the instrument is "closed".

Further properties are described in the annexes:

- Description T10159 revision 7;
- Documentation folder T10159-5.

Valid until

7 January 2029

Initially issued

7 January 2009

Remark

This revision replaces the previous versions, except for its documentation
folder.

Issuing Authority

NMi Certin B.V., Notified Body number 0122
7 April 2025

Certification Board

NMi Certin B.V.
Thijssseweg 11
2629 JA Delft
The Netherlands
T +31 88 636 2332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that
no liability is accepted and that the
manufacturer shall indemnify third-party
liability.

The designation of NMi Certin B.V. as Notified
Body can be verified at [http://](http://ec.europa.eu/growth/tools-databases/nando/)
ec.europa.eu/growth/tools-databases/nando/

Reproduction of the complete
document only is permitted.

This document is digitally signed
and sealed. The digital signature
can be verified in the blue ribbon at
the top of the electronic version of
this certificate.