


Issued by NMI Certin B.V.,
designated and notified by the Netherlands to perform tasks with respect to conformity assessment procedures mentioned in article 17 of Directive 2014/32/EU, after having established that the measuring instrument meets the applicable requirements of Directive 2014/32/EU, to:

Manufacturer MeteRSit S.r.l.
Viale dell'Industria, 34
35129 Padova
Italy

Measuring instrument **Thermal-mass flow gas meter**
Type : x4850xx, x4851xx
x4852xx, x4853xx
x4854xx, x4855xx
x4858xx,
NSM G4, NSM G6

Manufacturer's mark or name : 

Destined for the measurement of : Gas volume of natural gas or hydrogen. See description §1.2.4
Accuracy class : Class 1,5
Environment classes : M1 / E2
Gas temperature range : -25 °C / +55 °C
Ambient temperature range : -25 °C / +55 °C
Designed for : Condensing humidity
Intended location : Open

Further properties are described in the annexes:
– Description T10362 revision 80;
– Documentation folder T10362-63.

Valid until 1 July 2031

Initially issued 1 July 2011

Remark This revision replaces the earlier versions, except for its documentation folder.

Issuing Authority **NMI Certin B.V., Notified Body number 0122**
20 March 2026

Certification Board

NMI Certin B.V.
Thijsseweg 11
2629 JA Delft
The Netherlands
T +31 88 636 2332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the manufacturer shall indemnify third-party liability.

The designation of NMI Certin B.V. as Notified Body can be verified at <http://ec.europa.eu/growth/tools-databases/nando/>

Reproduction of the complete document only is permitted.

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon at the top of the electronic version of this certificate.