

Issued by NMI Certin B.V.,
designated and notified by the Netherlands to perform tasks with respect to
conformity assessment procedures mentioned in article 17 of Directive
2014/32/EU, after having established that the measuring instrument meets
the applicable requirements of Directive 2014/32/EU, to:

Manufacturer Flow Meter Group B.V.
Innovatieweg 32
7007 CD Doetinchem
The Netherlands

Measuring instrument **A Rotary Displacement Gas Meter**
Type : FMR
FMR-Dual
FMR-HTB

Manufacturer's mark or name : Flow Meter Group B.V.

Destined for the measurement of : Gas volume
Accuracy class : Class 1,0 or 1,5 (see tables under
1.2.1 of the description)

Environment classes : M1 / E2
Temperature range (FMR/FMR-Dual) : -40 °C / +70 °C (also see chapter
1.1.2)

Temperature range (FMR-HTB) : -10 °C / +70 °C

Further properties are described in the annexes:

- Description T10372 revision 15;
- Documentation folder T10372-11;
- Evaluation report ER10372 revision 15.

Valid until 30 August 2031

Initially issued 30 August 2011

Remark This revision replaces the earlier versions, except for its documentation
folder.

Issuing Authority **NMI Certin B.V., Notified Body number 0122**
6 August 2026

Certification Board

NMI Certin B.V.
Thijssseweg 11
2629 JA Delft
The Netherlands
T +31 88 636 2332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision
that no liability is accepted and that the
manufacturer shall indemnify third-party
liability.

The designation of NMI Certin B.V. as
Notified Body can be verified at [http://
ec.europa.eu/growth/tools-databases/nando/](http://ec.europa.eu/growth/tools-databases/nando/)

Reproduction of the complete
document only is permitted.

This document is digitally signed and
sealed. The digital signature can be
verified in the blue ribbon at the top
of the electronic version of this
certificate.