



# EU-type examination certificate

Number **T10377** revision 10  
Project number 3870694  
Page 1 of 1

Issued by

NMi Certin B.V.,  
designated and notified by the Netherlands to perform tasks with respect to  
conformity assessment procedures mentioned in article 17 of Directive  
2014/32/EU, after having established that the measuring instrument meets  
the applicable requirements of Directive 2014/32/EU, to:

Manufacturer

Zhejiang Cangnan Instrument Group Co., Ltd.  
No. 345, Strait Avenue, Lingxi Town, Cangnan County  
Wenzhou City, Zhejiang Province  
P.R. China 325800

Measuring instrument

**A Rotary Displacement Gas Meter**

Type : RM Series

Manufacturer's mark or name : Zhejiang Cangnan Instrument Group  
Co., Ltd.

Destined for the measurement of : Gas volume

Accuracy class : Class 1,0

Environment classes : M1

Temperature range : Depending on  $Q_{min}$ , see table 1.2.1

Maximum pressure : 20 bar(g)

Further properties are described in the annexes:

- Description T10377 revision 10;
- Documentation folder T10377-6.

Valid until

7 October 2031

Initially issued

7 October 2011

Remark

This revision replaces the earlier versions, except for its documentation  
folder.

Issuing Authority

**NMi Certin B.V., Notified Body number 0122**  
8 May 2025

Certification Board

**NMi Certin B.V.**  
Thijsseweg 11  
2629 JA Delft  
The Netherlands  
T +31 88 636 2332  
[certin@nmi.nl](mailto:certin@nmi.nl)  
[www.nmi.nl](http://www.nmi.nl)

This document is issued under the provision that  
no liability is accepted and that the  
manufacturer shall indemnify third-party  
liability.

The designation of NMi Certin B.V. as Notified  
Body can be verified at [http://](http://ec.europa.eu/growth/tools-databases/nando/)  
[ec.europa.eu/growth/tools-databases/nando/](http://ec.europa.eu/growth/tools-databases/nando/)

Reproduction of the complete  
document only is permitted.

This document is digitally signed  
and sealed. The digital signature  
can be verified in the blue ribbon at  
the top of the electronic version of  
this certificate.

