	<b>M</b>	E U - t	examination certificate	
(				Number <b>T10617</b> revision 6 Project number 3924779 Page 1 of 1
	Issued by	NMi Certin B.V., designated and notified by the Netherlands to perform tasks with respect to conformity assessment procedures mentioned in article 17 of Directive 2014/32/EU, after having established that the measuring instrument meets the applicable requirements of Directive 2014/32/EU, to:		
	Manufacturer	Goldcard Smart Group Co., Ltd. No. 158 Jinqiao Street Qiantang District, Hangzhou Zhejiang 310018 P.R. China		
	Measuring instrument	<b>Diaphragm gas meter</b> Type		JK/Gxx, JGMLxxF, JGDxx-SF, JGDxx-NF, JGDxx-NB, JGDxx-LR (xx is 1.6, 2.5 or 2.5W)
		Manufacturer's mark or name :		Goldcard Smart Group Co., Ltd.
		Destined for the measureme Accuracy class Environment classes Temperature range Location	:	Gas volume Class 1,5 M1 / E1 - 25 °C / +55 °C Closed
		Further properties are described in the annexes: – Description T10617 revision 6; – Documentation folder T10617-4.		
	Valid until	16 March 2035	March 2035	
	Initially issued	16 March 2015		
	Remark	This revision replaces the ea folder.	rlier version	s, excluding its documentation



## NMi Certin B.V., Notified Body number 0122 11 April 2025

## **Certification Board**

This document is issued under the provision that no liability is accepted and that the manufacturer shall indemnify third-party liability.

The designation of NMi Certin B.V. as Notified Body can be verified at http:// ec.europa.eu/growth/tools-databases/nando/

t Reproduction of the complete document only is permitted.

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon at the top of the electronic version of this certificate.



NMi Certin B.V. Thijsseweg 11 2629 JA Delft The Netherlands T +31 88 636 2332 certin@nmi.nl www.nmi.nl