



EU - type examination certificate

Number **T11664** revision 6
Project number 3983639
Page 1 of 1

Issued by

NMi Certin B.V.,
designated and notified by the Netherlands to perform tasks with respect to
conformity modules mentioned in article 17 of Directive 2014/32/EU, after
having established that the Measuring instrument meets the applicable
requirements of Directive 2014/32/EU, to:

Manufacturer

KROHNE Altometer
Kerkeplaat 12
3313 LC Dordrecht
The Netherlands

Measuring instrument An **Ultrasonic Gas Meter**
Type

: ALTOSONIC V12
ALTOSONIC V12 Check
ALTOSONIC V12D
ALTOSONIC V12D Check
ALTOSONIC V12E
ALTOSONIC V12E Check

Manufacturer's mark or name : KROHNE
Destined for the measurement of : Gas volume
Accuracy class : Class 0,5 – V12E/V12
Class 1,0 – V12D/V12
Environment classes : M2 / E2
Ambient temperature range : -40 °C / +55 °C
Gas temperature range : See description §1.2.5
Designed for : Condensing humidity
Intended location : Open

Further properties are described in the annexes:

- Description T11664 revision 6;
- Documentation folder T11664-6.

Valid until

3 June 2029

Initially issued

3 June 2019

Remark

This revision replaces the previous versions, including the documentation
folder.

Issuing Authority

NMi Certin B.V., Notified Body number 0122
21 November 2025

Certification Board

NMi Certin B.V.
Thijssseweg 11
2629 JA Delft
The Netherlands
T +31 88 636 2332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that
no liability is accepted and that the
manufacturer shall indemnify third-party
liability.

The designation of NMi Certin B.V. as Notified
Body can be verified at [http://
ec.europa.eu/growth/tools-databases/nando/](http://ec.europa.eu/growth/tools-databases/nando/)

Reproduction of the complete
document only is permitted.

This document is digitally signed
and sealed. The digital signature
can be verified in the blue ribbon at
the top of the electronic version of
this certificate.

