



EU-type examination certificate

Number **T13113** revision 0
Project number 4015507
Page 1 of 1

Issued by

NMi Certin B.V.,
designated and notified by the Netherlands to perform tasks with respect to
conformity assessment procedures mentioned in article 17 of Directive
2014/32/EU, after having established that the measuring instrument meets
the applicable requirements of Directive 2014/32/EU, to:

Manufacturer

RMG Messtechnik GmbH
Otto-Hahn-Straße 5
35510, Butzbach
Germany

Measuring instrument

A Rotary Displacement Gas Meter

Type	: RRM 100
Manufacturer's mark or name	: RMG
Destined for the measurement of	: Gas volume
Accuracy class	: Class 1,0
Environmental classes	: M2 / E1
Gas and ambient temperature range	: -25 °C / +55 °C

Further properties are described in the annexes:

- Description T13113 revision 0;
- Documentation folder T13113-1.

Initially issued

7 November 2025

Valid until

7 November 2035

Issuing Authority

NMi Certin B.V., Notified Body number 0122
7 November 2025

Certification Board

NMi Certin B.V.
Thijssseweg 11
2629 JA Delft
The Netherlands
T +31 88 636 2332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that
no liability is accepted and that the
manufacturer shall indemnify third-party
liability.

The designation of NMi Certin B.V. as Notified
Body can be verified at [http://](http://ec.europa.eu/growth/tools-databases/nando/)
ec.europa.eu/growth/tools-databases/nando/

Reproduction of the complete
document only is permitted.

This document is digitally signed
and sealed. The digital signature
can be verified in the blue ribbon at
the top of the electronic version of
this certificate.

