

EU-type examination certificate

Number **T10628** revision 15
Project number 3928235
Page 1 of 1

Issued by

NMi Certin B.V.,
designated and notified by the Netherlands to perform tasks with respect
to conformity assessment procedures mentioned in article 17 of Directive
2014/32/EU, after having established that the measuring instrument meets
the applicable requirements of Directive 2014/32/EU, to:

Manufacturer

METER&CONTROL d.o.o.
Veljka Dugoševića 54
11050 Belgrade
Serbia

Measuring instrument A static **Active Electrical Energy Meter**

Type : SM40xD
Manufacturer's mark or name : METER&CONTROL d.o.o.
Reference voltage : 230 V
Reference current : 5 A
Destined for the measurement of : electrical energy, in a
- single-phase two-wire network
Accuracy class : A or B
Environment classes : M1 / E2
Temperature range : -40 °C / +70 °C

Further properties are described in the annexes:

- Description T10628 revision 15;
- Documentation folder T10628-9.

Initially issued

18 September 2015

Valid until

18 September 2035

Remark

This revision replaces the earlier versions, except for its documentation
folder.

Issuing Authority

NMi Certin B.V., Notified Body number 0122
18 September 2025

Certification Board