



EU-type examination certificate

Number **T10806** revision 8
Project number 3984836
Page 1 of 1

| | |
|----------------------|---|
| Issued by | NMi Certin B.V., designated and notified by the Netherlands to perform tasks with respect to conformity modules mentioned in article 17 of Directive 2014/32/EU, after having established that the Measuring instrument meets the applicable requirements of Directive 2014/32/EU, to: |
| Manufacturer | Shenzhen Kaifa Technology (Chengdu) Co., Ltd. No. 99 Tianquan Rd. Hi-Tech Development Zone Chengdu, (PC: 611731) P.R. China |
| Measuring instrument | A static Active Electrical Energy Meter Type : MA304H3, MA304H4, MA304H3E, MA304H4D as direct connected meter (DCM) MA304T3, MA304T4 as transformer connected meter (TCM) Manufacturer's mark or name : KAIFA Reference voltage : 3x230/400 V or 3x230 V Reference current : 5 A (DCM) 1 A (TCM) Destined for the measurement : electrical energy, in a of - three-phase four-wire network - two-phase three-wire network Accuracy class : B or A (for DCM) C or B (for TCM) Environment classes : M2 / E2 Temperature range : -40 °C / +70 °C Further properties are described in the annexes: – Description T10806 revision 8; – Documentation folder T10806-7. Valid until 26 August 2026 Initially issued 26 August 2016 Remarks This revision replaces the earlier version(s), including its documentation folder. |

Issuing Authority **NMi Certin B.V., Notified Body number 0122**
05 September 2025

Certification Board

NMi Certin B.V.
Thijsseweg 11
2629 JA Delft
The Netherlands
T +31 88 636 2332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision
that no liability is accepted and that the
manufacturer shall indemnify third-party
liability.

The designation of NMi Certin B.V. as Notified
Body can be verified at [http://
ec.europa.eu/growth/tools-databases/nando/](http://ec.europa.eu/growth/tools-databases/nando/)

Reproduction of the complete
document only is permitted.

This document is digitally signed
and sealed. The digital signature
can be verified in the blue ribbon at
the top of the electronic version of
this certificate.

