

EU-type examination certificate

Number **T11968** revision 12
Project number 3976972
Page 1 of 1

Issued by

NMi Certin B.V.,
designated and notified by the Netherlands to perform tasks with respect
to conformity assessment procedures mentioned in article 17 of Directive
2014/32/EU, after having established that the measuring instrument meets
the applicable requirements of Directive 2014/32/EU, to:

Manufacturer

Landis+Gyr
78th km National Road Athens-Corinth
Corinth GR-201 00
Greece

Authorised
representative

Landis+Gyr Oy
Paperitehtaankatu 9
FI-40100 Jyväskylä
Finland

Measuring instrument A static **Active Electrical Energy Meter**

Type : E450 ZMXi3x0xQ... S5

Manufacturer's mark or name : Landis+Gyr

Reference voltage : 3x230/400V

Reference current : 5 A or 10 A

Destined for the measurement of : electrical energy, in a
- three-phase four-wire network
- two-phase three-wire network
- one-phase two-wire network

Accuracy class : A or B

Environment classes : M2 / E2

Temperature range : -40 °C / +70 °C

Further properties are described in the annexes:

- Description T11968 revision 12;
- Documentation folder T11968-8.

Valid until

27 October 2030

Initially issued

27 October 2020

Remarks

This revision replaces the earlier versions, including its documentation
folder.

Issuing Authority

NMi Certin B.V., Notified Body number 0122
21 August 2025

Certification Board