

EU-type examination certificate

Number **T12420** revision 12
Project number 4020292
Page 1 of 1

Issued by

NMi Certin B.V.,
designated and notified by the Netherlands to perform tasks with respect to
conformity assessment procedures mentioned in article 17 of Directive
2014/32/EU, after having established that the measuring instrument meets
the applicable requirements of Directive 2014/32/EU, to:

Manufacturer

Landis+Gyr
78 km National Road Athens
Corinth GR-20100
Greece

Authorised
representative

Landis+Gyr Oy
Paperitehtaankatu 9
FI-40100 Jyväskylä
Finland

Measuring instrument

A static **Single Phase Active Electrical Energy Meter**

Type : E360-AD2D...S2 or E360-AB3D...S2

Manufacturer's mark or name : Landis+Gyr

Reference voltage : 230 V

Reference current : 5 A or 10 A

Destined for the measurement of : electrical energy, in a
- single-phase two-wire network

Accuracy class : A or B

Environment classes : M1 / E2

Temperature range : -40 °C / +70 °C

Further properties are described in the annexes
- Description T12420 revision 12
- Documentation folder T12420-6

Valid until

4 January 2033

Initially issued

4 January 2023

Remarks

This revision replaces the earlier versions, except for its documentation
folder.

Issuing Authority

NMi Certin B.V., Notified Body number 0122
13 October 2025

Certification Board