

EU-type examination certificate

Number **T12814** revision 1
Project number 3939260
Page 1 of 1

Issued by

NMi Certin B.V.,
designated and notified by the Netherlands to perform tasks with respect to
conformity assessment procedures mentioned in article 17 of Directive
2014/32/EU, after having established that the measuring instrument meets
the applicable requirements of Directive 2014/32/EU, to:

Manufacturer

Enphase Energy Inc.
47281 Bayside Parkway
Fremont, CA 94538
United States of America

Authorised
representative

Enphase Energy NL B.V.
Het Zuiderkruis 65
5215 MV 's-Hertogenbosch
The Netherlands

Measuring instrument **A static **Active Electrical Energy Meter integrated in an EV Charging
System****

Type : IQ EV Charger 2
Manufacturer's mark or name : Enphase Energy Inc.
Reference voltage : 3x230/400 V, 3x230 V, 230 V
Reference current : 5 A

Destined for the measurement of : electrical energy, in a
- three-phase four-wire network
- three-phase three-wire network
- single-phase two-wire network

Accuracy class : B or A
Environment classes : M1 / E1
Temperature range : -40 °C / +55 °C

Further properties are described in the annexes:
- Description T12814 revision 1;
- Documentation folder T12814-1.

Valid until

18 November 2034

Initially issued

18 November 2024

Remark

This version replaces the earlier version(s) except for the documentation
folder.

Issuing Authority

NMi Certin B.V., Notified Body number 0122
18 March 2025

Certification Board