

EU-type examination certificate

Number **T13110** revision 1
Project number 4069617
Page 1 of 1

Issued by

NMi Certin B.V.,
designated and notified by the Netherlands to perform tasks with respect to
conformity assessment procedures mentioned in article 17 of Directive
2014/32/EU, after having established that the measuring instrument meets
the applicable requirements of Directive 2014/32/EU, to:

Manufacturer

Shenzen Kaifa Technology (Chengdu) Co. Ltd.
No. 99 Tianquan Rd.
Hi-Tech Development Zone
611731 Chengdu
P.R. China

Measuring instrument

A static **Active Electrical Energy Meter**

Type : MA309H4SDE

Manufacturer's mark or name : Shenzen Kaifa Technology (Chengdu)
Co. Ltd

Reference voltage : 230V or 3x230/400 V

Reference current : 5 A

Destined for the measurement of : electrical energy, in a
- three-phase four-wire network
- single-phase two-wire network

Accuracy class : A or B

Environment classes : M1 / E2
M2 / E1

Temperature range : -40 °C / +70 °C

Further properties are described in the annexes:

- Description T13110 revision 1;
- Documentation folder T13110-1.

Valid until

29 December 2035

Initially issued

29 December 2025

Remarks

This revision replaces the earlier version, except for its documentation folder.

Issuing Authority

NMi Certin B.V., Notified Body number 0122
27 January 2026

Certification Board