

## EC-type examination certificate

Number **T10125** revision 0 Project number 705904 Page 1 of 1

Issued by

NMi Certin B.V.

Hugo de Grootplein 1 3314 EG Dordrecht The Netherlands

Notified Body number 0122

In accordance with

The Metrologiewet (Stb. 2006, 137) as Dutch implementation of Directive 2004/22/EC

on measuring instruments (MID).

Manufacturer

Seltec S.r.l. Via di Prato 74 50041 Calenzo

Italy

In respect of

A model of an Exhaust Gas Analyzer.

Type

: V-Gas

Characteristics

Accuracy class 0

Electromagnetic environment class E2 Temperature range +5 °C / +40 °C

In the description number T10125 revision 0 further characteristics are described.

Valid until

21 March 2018

Description and The instrument is described in the description number T10125 revision 0 documentation and documented in the documentation folder T10125-1, appertaining to this EC-type examination certificate.

Dordrecht, 21 March 2008

NMi Certin B.V.

Ing. C. Oosterman

Manager Product Certification

