

## EC type-approval certificate

Number **T7046** revision 1 Project number 606437 Page 1 of 5

Issued by NMi Certin B.V.

Hugo de Grootplein 1 3314 EG Dordrecht The Netherlands

In accordance

The Council Directive 90/384/EEC on non-automatic weighing instruments.

Manufacturer

A&D Instruments Ltd. 24 Blacklands Way Abingdon Business Park Abingdon, Oxfordshire OX14 1DY United Kingdom

In respect of

A class  $\bigcirc$  and  $\bigcirc$ , electronic, single-interval, **non-automatic weighing** 

instrument.

Type

: GX-8K / GX-10K / GX-12K / GX-20K / GX-30K

## Characteristics

+ + + + + + + + + +	+ + + + + 10 + + + + +
8.1 kg ≤ Max ≤ 10.1 kg	12 kg ≤ Max ≤ 31 kg
e = 0.1 g	e = 1 g
d = 0.01 g	+ + + + + d = 0.1 g + + + + +
n ≤ 10100 divisions	n ≤ 31000 divisions

Operating temperature range: +10 °C / +30 °C

In the description number T7046 revision 1 further characteristics are described.

Valid until 4 30 November 2016

Description and The instrument is described in the description number T7046 revision 1 and documentation documented in the documentation folder T7046-1, appertaining to this

EC type-approval certificate.

Remarks This version replaces the earlier version, except for its documentation folder.

The Notified Body no. 0122 NMi Certin, 13 August 2009

C. Oosterman

**Head Certification Board** 

NMi Certin BV Hugo de Grootplein 1 3314 EG Dordrecht PO Box 394 3300 AJ Dordrecht, NL T +31 78 6332332 F +31 78 6332309 certin@nmi.nl Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMi (see "Regulation objection and appeal against decisions of NMi")

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

Reproduction of the complete document is permitted.

