

EC type-approval certificate

Number T7401 revision 0 Project number 801792 Page 1 of 6

Issued by

NMi Certin B.V.

Hugo de Grootplein 1 3314 EG Dordrecht The Netherlands

Notified Body Number 0122

In accordance

with

The Council Directive 90/384/EEC on non-automatic weighing instruments.

Applicant

Micron s.r.l.

Via Ticino, 53

21044 Cavaria con Premezzo

In respect of

A class (III) electronic, single- or multi-interval or multi-range, non-automatic weighing instrument, intended to be used for direct sales to the public.

Manufacturer

: Micron, s.r.l.

Type

MR

Characteristics $3 \text{ kg} \leq \text{Max} \leq 30 \text{ kg}$

e ≥ 1 q

n ≤ 6000 divisions for single interval and multi-range instruments;

 $n \le 3000$ divisions per partial weighing range for multi-interval instruments,

with a maximum of two weighing ranges

In the description number T7401 revision 0 further characteristics are described.

Valid until

8 May 2018

Description and The instrument is described in the description number T7401 revision 0 and documentation documented in the documentation folder T7401-1, appertaining to this

EC type-approval certificate.

Dordrecht, 8 May 2008

NMi Certin B.V.

Ing. C. Oosterman

Manager Product Certification

Nederlands Meetinstituut Hugo de Grootplein 1 3314 EG Dordrecht Telephone +31 78 6332332

Telefax +31 78 6332309

NMi B.V.

(Chamber of Commerce no.27.228.701)

Subsidiary companies:

NMi Van Swinden Laboratorium B.V. (27228703) NMi Certin B.V. (27.233.418) Verispect B.V. (27.228.700)

This document is issued under the provision that NMi. B.V. nor its subsidiary companies accept any liability.

Reproduction of the complete document is allowed. Parts of the document may only be reproduced after written permission.